



MINISTRY OF HEALTH
NATIONAL INSTITUTE FOR FOOD CONTROL (NIFC)

Address: No 65 Pham Than Duat Street, Mai dich Ward, Cau Giay District, Hanoi City, Vietnam.
Phone: 84-2432262216; Fax: 84-2439335718; Website: www.nifc.gov.vn

No: 6328/PKN-VKNQG

TEST REPORT

1. Name of sample: ANTI-FOG MEDICAL MASK
2. Sample code: 04202596/DV.1
3. Sample description: Sample is put in a plastic bag, 10 pieces/bag - Quantity: 01
MFG - EXP: No information; No storage sample
4. Number of sample: 01 sample
5. Sample storage time: No storage
6. Sample received date: May 16, 2020
7. Testing time: March 22, 2020 - March 28, 2020.
8. Customer: ORANGE PHARMACEUTICAL COMPANY LIMITED
No 10, Lane 5, Cao Thang Street, Yet Kieu Ward, Ha Dong District
Hanoi City, Vietnam.
9. Test result(s): Microbiological test.

No	Parameter	Unit	Test Method	Result
9.1	Bacterial filtration efficiency (BFE)	/piece	Ref. TCVN 8389-2:2010 & BS EN 14683: 2019	98,66 %

Note: Bacterial filtration efficiency is equivalent to the protective ability against bacterial infection of the face mask.

Hanoi, March 28, 2020

On behalf of Director General
Head of Quality Assurance Laboratory



Le Thi Phuong Thao, PhD

- 1 The test results stated in this sample are valid for the sample
2. No part of the test results may be quoted without the NIFC's written consent.
3. Information about the sample and the customer is recorded at the request of the place where the sample was sent